	Application No.	Applicant(s)	
Notice of Allowability	09/931,804	MANN ET AL.	
	Examiner	Art Unit	
	Kennedy Schaetzle	3762	
The MAILING DATE of this communication apperall claims being allowable, PROSECUTION ON THE MERITS IS herewith (or previously mailed), a Notice of Allowance (PTOL-85) NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RI	(OR REMAINS) CLOSED in this app or other appropriate communication GHTS. This application is subject to	olication. If not include will be mailed in due of	d course. THIS
1. \boxtimes This communication is responsive to the amendment received	ved 1/13/05.		
2. \boxtimes The allowed claim(s) is/are <u>31-33 and 36-53</u> .			
3. \boxtimes The drawings filed on <u>04 December 2001</u> are accepted by	the Examiner.		
4. Acknowledgment is made of a claim for foreign priority una All b) Some* c) None No	e been received. been received in Application No cuments have been received in this is of this communication to file a reply dENT of this application. itted. Note the attached EXAMINER bes reason(s) why the oath or declara of the submitted. Son's Patent Drawing Review (PTO- s Amendment / Comment or in the Co .84(c)) should be written on the drawing he header according to 37 CFR 1.121(c) sit of BIOLOGICAL MATERIAL re	national stage applicated of the requirement of the	uirements OTICE OF
Attachment(s) 1. Notice of References Cited (PTO-892) 2. Notice of Draftperson's Patent Drawing Review (PTO-948) 3. Information Disclosure Statements (PTO-1449 or PTO/SB/O Paper No./Mail Date 4. Examiner's Comment Regarding Requirement for Deposit of Biological Material	5. Notice of Informal P 6. Interview Summary Paper No./Mail Dai 7. Examiner's Amendr 8. Examiner's Stateme 9. Other	(PTO-413), te nent/Comment	·

Application/Control Number: 09/931,804 Page 2

Art Unit: 3762

EXAMINER'S AMENDMENT

1. An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

The application has been amended as follows:

TITLE

The title was changed to: "Implantable Stimulator Methods for Treatment of Incontinence and Pain" to better reflect the claimed invention.

CLAIMS

Claims 2, 3, 5-19 and 21-30 were canceled as being drawn to a nonelected invention, with no remaining allowable generic claims.

Reasons for Allowance

2. The following is an examiner's statement of reasons for allowance: the prior art of record fails to disclose a method of treating patients comprising the steps of delivering a stimulation pulse to the pudendal nerve by perineally or vaginally implanting electrodes and/or a catheter discharge portion adjacent to the pudendal nerve and by implanting the stimulator at a location remote from the pudendal nerve. The examiner agrees with the applicants' statements regarding the Lu et al. reference and its teaching to stimulate the cavernous nerve —a nerve distinct from the pudendal nerve. Furthermore, it is unclear that a stimulator designed for stimulation of the cavernous nerve with electrodes wrapped about said nerve, would be capable of stimulating the pudendal nerve since electrode location would have a distinct impact on stimulation parameters. Nor would it appear that one would desire the inadvertent stimulation of nerves such as the

Art Unit: 3762

pudendal when attempting to stimulate the cavernous because of the unwanted sideeffects of such stimulation. One for example would not wish to stimulate penile erection when treating incontinence.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Conclusion

3. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Kennedy Schaetzle whose telephone number is 571 272-4954. The examiner can normally be reached on M-W and F from 9:30 -6:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Angela Sykes can be reached on M-F at 571 272-4955. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

KJS March 31, 2005